

Edition 04 / 2011

HIOSSEN Inc. www.hiossen.com 85 Ben Fairless Dr, Fairless Hills, PA 19030, USA TEL : 1-888-678-0001 FAX : 1-267-759-7004



Lateral Approach - Sinus KIT

LASK



 \oplus

Lateral Approach - Sinus KIT

LASK (Lateral Approach - Sinus KIT)

Contents

- 1) Introduction
- 2) Features
- 3) Components
- 4) Clinical indication and Case
- 5) Instruction for Use

Introduction

• Hiossen's Lateral Approach Sinus Kit is specifically designed for a fast and safe lateral approach to lift sinus membrane.

LASK offers various usage depending on the oral anatomy and surgical plans. LASK contains Dome drill to create a lateral window, Wide dome drill to widen the window, Core drill that can create core bone lid.

- · Dome and Core Drills to create lateral window.
- · Stopper system to prevent excessive and over drilling.
- · Dome Drill Exceptional cutting ability through the combined use of macro and micro blades.
- \cdot Core Drill continued successful design concept of CAS Drills
- \cdot Wide dome drill and side wall drill to enlarge the window









Side window expansion

Dome & Core drill

Stopper system

Maximized cutting efficiency

HÍÓSSEN LASK

LASK (Lateral Approach - Sinus KIT)

Hiossen's Lateral Approach Sinus Kit is specifically designed for a fast and safe approach to lateral sinus lifts.

Features 🗌 • Dome and Core Drills provide ideal approaches to the opening of the lateral wall. <Dome drill> <Core drill> (Case by Dr. D.H. Lee) (Case by Dr. Y.S. Cho) • Excellent cutting ability of the Dome and Core Drills Company B Company A NIP OSSTEM

3.5

2.5 3.0 CAS-Drill Concept

• Widen the window with Side Wall Drill



The blade design of LAS Drill follows the successful design concept of CAS Drill (Crestal Approach Sinus Kit)

 Wide Dome Drill and Side Wall Drill used to enlarge the window

0.0

0.5 1.0 1.5 2.0

Combination of macro and micro blades offer excellent cutting ability

Micro

Macro





• Hiossen's unique stopper system for depth control. (a total of 6 stoppers : 0.5 / 1.0 / 1.5 / 2.0 / 2.5 / 3.0mm)









HIÓSSEN LASK

LASK comes with Dome Drill and Core Drill that provide various approaches to a fast and safe lift of sinus membrane.

Safe Elevation of Sinus Membrane

<Dome drill>

Minimizing direct contact with the membrane by forming a bone lid



Formation of bone particles between the cutting blades



Effective depth control by stopper system (0.5mm increment)Stopper can prevent soft tissue damage



<Core drill>

Round-shaped cutting edge minimizes direct contact with the membrane



· Formation of bone particles between the cutting blades



Effective depth control by stopper system (0.5mm increment)Stopper can prevent soft tissue damage




Ease of Use

<Dome drill>

• The blade head can be perpendicular to the bone to perform an osteotomy.




For limited space at surgical site, Dome Drill can be tilted to drill.






<Core drill>



Components

• Dome drill Ø 5.5/Ø 7.0 & Wide dome drill Ø 7.0, Core drill Ø 5.5/Ø 7.0, Side wall drill, Bone separator, Stopper 0.5 / 1.0 / 1.5 / 2.0 / 2.5 / 3.0 * Sinus Kit sold separately.

LASK




HÍÓSSEN LASK

Clinical indication and Case 1) Dome drill, Side wall drill for widening the window by Dr. D.H. Lee Jse of Side Wall Drill to n be tilted fo tight spaces widen the window 2) Dome drill with stopper by Dr.Y.S. Cho Collecting of Drill Stopper autogenous bones by Dr.Y.S. Cho 3) Overlapped drilling to create enlarged window Window closing with the Overlapped drilling to the bone reate enlarged window bone lids 4) Overlapped drilling to create enlarged window by Dr. D.H. Lee Overlapped drilling to Membrane safely lifted Use of stopper to control depth reate enlarged windo Ø 7.0 Ø 7.0 5) Overlapped drilling to create enlarged window with two different drill sizes by Prof. J.C. Jeong Create the bone lid Overlapped drilling with two different drill sizes Ø 5.5 70 6) Combined use of Core Drill and CAS Drill by Dr. M.S. Kim Capturing bone lid IS D with core drill by Dr. K.D. Jeong 7) Combined use of Core Drill and Hydraulic Lifter from CAS Kit Using Hydraulic lift Lifted Membrane Use of Core F to lift mem

Instruction for Use

LASK Care & Maintenance



- ① Prepare tools for surgery by soaking them in a "saline solution" or in "distilled water."
- ② After surgery: All tools should be soaked in an "alcohol solution".
 - Avoid using Hydrogen Peroxide.

- Hydrogen Peroxide will discolor laser markings and anodized surfaces.

- ③ Tools should be cleaned thoroughly with distilled or tap water to wash away any remaining blood and foreign material.
- ④ Completely dry all tools using a dry cloth or warm air.
- In the KIT case.In the color coding when placing the tools back in the case)
- ⑥ After placing all the tools back into the kit, dry the entire kit in an Autoclave (132° for 15 minutes) and then store the kit at room temperature.

NOTES:

It is recommended to re-sterilize the surgical KIT right before surgery. (132°c; for 15 minutes) Immediately after surgery, all the tools should be cleaned and stored. The LASK has a one year warranty on all parts & case. The recommended usage of the drills is 50 times.